

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/23/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.86: 0.079

LOT #: 02511080A2

EXPIRATION: 02/01/2013

BLANK TEST	0.000	14:09
INTERNAL STANDARD	VERIFIED	14:10
EXTERNAL STANDARD	0.080	14:10
BLANK TEST	0.000	14:11
EXTERNAL STANDARD	0.080	14:11
BLANK TEST	0.000	14:12
EXTERNAL STANDARD	0.079	14:12
BLANK TEST	0.000	14:13
EXTERNAL STANDARD	0.079	14:13
BLANK TEST	0.000	14:14
EXTERNAL STANDARD	0.079	14:14
BLANK TEST	0.000	14:15

Average = 0.0794

Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 1.01

PIC: 3.02

Modem: 2.0

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C

PASSED

Breath Tube = 43.7°C

PASSED

PUMP INFO

Flow Rate = 5.191 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 20 11

Notary Public's Signature

My Commission Expires on with Office

Notary Name



CRF
7/11/11